



KAROLINSKA
Universitetssjukhuset

EQUALIS

Technical Report on the Global HPV LabNet DNA Screening Proficiency Panel 2022

**Carina Eklund¹, Camilla Lagheden¹, Emel Yilmaz¹, Sara Arroyo
Mühr¹, Marina Lilja², Ola Forslund³ & Joakim Dillner¹**

1 HPV International Reference Center, Department of Clinical Science, Intervention and Technology, Karolinska Institute and Karolinska University Hospital, Stockholm, Sweden

2 Equalis, Uppsala, Sweden

3 Clinical Microbiology, Skåne Regional and University Laboratories, Lund Sweden

Table of Contents

1. Introduction.....5

2. Aims.....5

3. Methods.....5

 3.1 Panel composition.....5

 3.2 Validation of the SP6

 3.3 Distribution of the PP6

 3.4 Data analysis.....8

4. Results9

 4.1 Results by assays used.....9

 4.2 Results analysed by assay9

 4.2.2 Results achieved by participating laboratories12

 4.3 Results by sample number.....14

 4.4 Analysis of false positive results15

5. Discussion.....17

6. Conclusions and recommendations19

Annex 1: Call for participation: Global HPV LabNet Proficiency Study for evaluating HPV DNA Typing Methods

Annex 2: Application form

Annex 3: Source of plasmids included

Annex 4: Participating laboratories

Tables

- Table 1: 2022 HPV Screening panel composition by randomised sample ID
- Table 2: Distribution of participating laboratories and submitted datasets in each WHO region.
- Table 3: Assays used for analyzing the HPV screening panel.
- Table 4: Proportion of data sets submitted by WHO region with ≥ 90 % proficient HPV typing results.
- Table 5: Overall proficiency of 2022 screening panel by assay type.
- Table 6: Number of datasets that have detected all HPV types in optional samples.
- Table 7: Proficiency of each sample. To achieve 100% proficiency full detection of HPV types 16 and 18 with 10 IU/ μ l, HPV types 31, 33, 45, 52 and 58 with 1000 genome equivalents/ μ l and no false positivity is required.
- Table 8: Number of false positive HPV types detected per data set reported by assay used.

Figures

- Figure 1: Global distribution of laboratories that submitted results for the 2022 HPV DNA SP.
- Figure 2: Type of assay in use for HPV DNA screening by WHO region.
- Figure 3: Proficiency for HPV DNA screening by WHO region. The figure includes 158 data set.



Abbreviations

AMR	Region of the Americas
EMR	Eastern Mediterranean Region
Equalis	External Quality Assessment in Laboratory Medicine in Sweden
EUR	European Region
GE	Genome Equivalent
GRL	Global Reference Laboratory
HPV	Human Papilloma Virus
HPV LabNet	HPV Laboratory Network
IU	International Unit
PP	Proficiency Panel
SEAR	South East Asian Region
SP	Screening Panel
WHO	World Health Organization
WPR	Western Pacific Region



Global HPV DNA screening proficiency panel 2022

Distributed in October 2022

1. Introduction

The elimination of cervical cancer is a globally prioritized health policy goal, formulated by the World Health Organization (WHO). An essential part of the elimination strategy is high coverage population-based cervical screening programs implemented worldwide with testing for the human papillomavirus (HPV), the main cause of cervical cancer. Evaluation of HPV testing assays, as actually applied in different laboratories, is fundamental to ensure the quality, reliability, and accuracy of HPV-based cervical screening programs.

A major method for achieving progress towards the elimination goal is development, preparation and validation of proficiency panels (PP) to qualify methods and laboratories. Therefore, we designed a proficiency panel tailored to assess the quality of HPV testing services used for cervical screening.

Call for participation in this screening proficiency study was sent to all laboratories that had participated in the HPV LabNet proficiency panels in previous years (Annex 1, 2).

2. Aims

The aims of this panel were:

1. To assess the proficiency of HPV screening assays when routinely used in laboratories worldwide.
2. To evaluate the sensitivity and type-specificity of HPV detection of the different HPV screening assays when routinely used in laboratories worldwide.
3. Identify problems with any assays routinely used.

3. Methods

3.1 Panel composition

Complete genomes of HPV cloned into plasmid vectors had been provided to the International HPV Reference Center by the respective proprietors with written approval for use in this SP. All samples were purified plasmids diluted in TE buffer (10 mM TRIS-HCl, 0.1 mM EDTA, pH 8.0) with 10



ng/μl of human placental DNA (Sigma-Aldrich no 7011). The HPV types included were: 16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, 59, 68a (HPV 68 prototype) and 68b (ME 180 isolate).

The composition of the panel is shown in Table 1.

Table 1: 2022 HPV Screening panel composition by randomised sample ID

HPV types	HPV genome equivalents (GE) or international unit (IU) (HPV 16, 18) per μl	Randomised Panel ID
16	1 IU	9
16	10 IU	2
18	1 IU	4
18	10 IU	7
31	1000 GE	1
33	1000 GE	8
45	1000 GE	11
52	1000 GE	3
58	1000 GE	12
31/33/45/52/58	100 ^a GE	5
35/39/51/56/59/68	100 ^a GE	10
Negative (TE buffer with 10 ng/μl human placenta DNA)	0	6

^a100 copies of each HPV type. IU: international units, GE: genome equivalents

3.2 Validation of the SP

The SP was pre-tested at GRL Sweden using a modified GP5+/6+ PCR followed by Luminex-based typing for HPV types 6, 11, 16, 18, 26, 30, 31, 33, 35, 39, 40, 42, 43, 45, 51, 52, 53, 54, 56, 58, 59, 61, 66, 67, 68a, 68b, 69, 70, 73, 74, 81, 82, 86, 87, 89, 90 and 91. As a second validation the SP was pre-tested using BD-COR Onclarity and Roche Cobas 4800.

3.3 Distribution of the PP

After pre-test validation by the International HPV Reference center in Sweden, the SP was compiled in September 2022 and distributed to laboratories throughout four WHO regions in October 2022, following the call for participation and requests received from the laboratories. The fee for participation was for commercial entities 330 Euros, whereas academic and public health entities had a fee of 140 Euros. Participants from low and lower-middle income countries could apply for waiving of the fee. As in previous years the HPV reference laboratory in Sweden prepared



the materials for the 2022 HPV DNA SP, but subcontracted the administration and distribution of the PP to Equalis (External quality assessment of laboratory medicine in Sweden; <https://www.equalis.se/en/> a public, non-profit company that administrates the external quality assessment for public health care laboratories in Sweden) that handled the logistics and distribution of the panel. This model has continued to work well and is a possible mode of operation of a sustainable long-term activity with global distribution of an HPV DNA typing proficiency panel. National Reference Laboratories, appointed by their government were invited to receive panels for free to organize a quality control program in their country. The reference laboratories in Norway, Germany and Argentina did order panels to organize local quality control programs.

One hundred fifty-eight datasets from 84 laboratories were obtained. The number (n) of laboratories submitting results per WHO Region is shown in Table 2. These are EUR (n = 29), SEAR (n = 2), WPR (n = 26) and AMR (n = 27). Fifty-three laboratories submitted a data set from one assay, eleven laboratories submitted data sets from 2 different assays, nineteen laboratories submitted data sets from 4 different assays and one laboratory submitted results from 7 different assays.

Table 2: Distribution of participating laboratories and submitted datasets in each WHO region.

WHO region	Participating laboratories (n)	Submitted datasets(n)
Total	84	158
AMR	27	35
EUR	29	36
SEAR	2	2
WPR	26	85




Figure 1: Global distribution of laboratories that submitted results for the 2022 HPV DNA SP.

3.4 Data analysis

Results analysed in this report include all results returned prior to the 5th of February 2022. Data was compiled by Equalis and transferred to the HPV reference laboratory in Sweden for analyses. Each data set was designated a number from 1 to 158. The data were analysed by region of the laboratory, by assay used and by HPV type.

From the data sets submitted, it was noted that participating laboratories used a range of commercial assays as well as 9 datasets submitted using in-house assays (Table 3). The proportion of correct HPV typing results, reported by the laboratory, was analyzed as data sets by laboratory and according to assay used.

This was the first screening proficiency panel, the requirements for proficiency were established by comparing data from the genotyping proficiency panel and a longitudinal cohort study with type-



specific quantitation of HPV in cervical samples taken up to 8.5 years before development of invasive cervical cancer.

A data set was considered proficient when it detected at least 10 international units (IU) / μl of HPV 16 and HPV 18, and 1000 genome equivalents (GE) / μl of HPV 31, 33, 45, 52 and 58. Two samples with HPV16 and 18 at 1 IU/ μl and a pool with 100 GE/ μl for HPV types 31, 33, 45, 52, 58 were also included but detection of these low virus amounts was not required for proficiency. In addition, the specificity of the reported types should be 100 % (no false positive result).

4. Results

Eighty-four participating laboratories submitted one hundred fifty-eight data sets.

Each data set submitted by each laboratory was analyzed and a feedback letter was sent to all participating laboratories that had paid the fee in February 2023.

4.1 Results by assays used

4.1.1 Commercial assays

A total of 149 data sets were obtained using commercially available tests. The most commonly used assays were the HybriBio Genotyping with Real-time PCR that was used in 21 laboratories. Two other HybriBio assays were used by 20 and 19 laboratories, Hybrid Capture 2 (Qiagen) was used in 19 laboratories (Table 3).

4.1.2 In-house assays

Nine of the data sets had been obtained using a variety of in-house assays (Table 3).

4.2 Results analysed by assay

4.2.1 Assay details

The different assays used for testing and typing of HPV as well as the number of submitted data sets and different part of the HPV genome targeted by each assay is shown in Table 3. The distribution of different assays in different WHO regions is shown in Figure 2.



Table 3: Assays used for analyzing the HPV screening panel.

HPV assay	Number of data sets	HPV region targeted (primers)
All assays	158	L1/L2/E1/E2/E4/E6/E7
Abbott/Abbott Alinity m (Abbott)	6	L1
Allplex HPV 28/Allplex HPV HR (Seegene)	2	
Ampfire HPV screen	2	E6 / E7
Anyplex II HPV28/Anyplex II HPV HR (Seegene)	8	
BD_Onclarity/Viper (BD)	7	E6 / E7
Cobas4800/6800 (Roche)	12	L1
Genotyping with Real-time PCR (HybriBio)	21	L1/L2/E1/E2/E4/E6/E7
Hybrid Capture 2 (Qiagen)	19	
14 High-risk (HybriBio)	19	E6 / E7
21 HPV GenoArray (HybriBio)	20	L1 (MY09)
37 HPV GenoArray (HybriBio)	19	L1 (MY09)
ScreenFire HPV RS	2	
TypeSeq/TypeSeq2	5	L1 / E6 / E7
Other commercial assays ^a	12	L1/L2/E1/E2/E4/E6/E7
Other In-house assays ^b	4	

^aIn the other commercial assays category, the following assay types are included: Line Probe assay, Screen +, Ampliquality HPV-type, HPV Screening real-time PCR, Alias: Harmonia HPV, Alias: Venus HPV, Cepheid Xpert HPV, NeoPlex, Sacace HPV screen and type, HPV Direct-Flow Chip, INNO-LiPA HPV and Acon Biotech High risk.

^bIn the other in-house assays category, the following assay types are included: multiplex qPCR, real-time PCR, MALDI-TOF, RFLP.




Figure 2: Type of assay in use for HPV DNA screening by WHO region.




Figure 3: Proficiency for HPV DNA screening by WHO region. The figure includes 158 data set.



Table 4: Proportion of data sets submitted by WHO region with ≥ 90 % proficient HPV typing results.

Region (data sets)	Proportion of laboratories with 100 % correct typing	Proportion of laboratories with ≥ 90 % correct typing
EUR (36)	81 %	83 %
SEAR (2)	50 %	50 %
AMR (35)	29 %	50 %
WPR (85)	96 %	96 %

4.2.2 Results achieved by participating laboratories

The criteria for proficiency was no false positive results. According to the criteria described in 3.4, 122 (77 %) data set out of the 158 data sets were 100 % proficient. All the datasets with HPV analysis results reported from BD Onclarity/Viper, Cobas4800/6800, Genotyping with Real-time PCR (HybriBio), HybriBio 14 High-risk, and HybriBio 21 HPV GenoArray showed 100% proficiency (Table 5).

The proportion of laboratory proficiency including all datasets grouped by WHO region is shown in Figure 3 and Table 4. The percent proficiency of detecting HPV types grouped by assay used for testing is shown in Table 5.

A total of 54/84 laboratories provided at least one fully proficient dataset. The performance of assays detecting HPV types at low concentrations where HPV detection was not required for proficiency is summarized in Table 6. The low concentration HPV18 sample (1 IU/ μ l) was detected by 80% of the datasets (126/158) and the low concentration HPV16 (1 IU/ μ l) was detected by 84% (133/158) of datasets. The low amounts of HPV 31/33/45/52/58 (100 GE/ul) and HPV 35/39/51/56/59/68 (100 GE/ul) were detected in 91% (143/158) and in 94% (148/158) of datasets, respectively (Table 6).



Table 5: Overall proficiency of 2022 screening panel by assay type.

HPV assay	Number of data sets	No. of proficient data sets				
		100 % proficient	99-90 % proficient	89-80 % proficient	<80 % proficient	Not proficient
All assays	158	122	11	15	9	1
Abbott/Abbott Alinity m (Abbott)	6	5	1	0	0	0
Allplex HPV 28/Allplex HPV HR (Seegene)	2	1	1	0	0	0
Ampfire HPV screen	2	1	1	0	0	0
Anyplex II HPV28/Anyplex II HPV HR ^a (Seegene)	8	5	2	1	0	0
BD_Onclarity/Viper (BD)	7	7	0	0	0	0
Cobas4800/6800 (Roche)	12	12	0	0	0	0
Genotyping with Real-time PCR	21	21	0	0	0	0
Hybrid Capture 2 (Qiagen)	19	0	0	13	6	0
14 High-risk (HybriBio)	19	19	0	0	0	0
21 HPV GenoArray (HybriBio)	20	20	0	0	0	0
37 HPV GenoArray (HybriBio)	19	18	1	0	0	0
ScreenFire HPV RS	2	0	2	0	0	0
TypeSeq/TypeSeq2	5	3	2	0	0	0
Other commercial assays ^a	12	9	0	1	2	0
Other In-house assays ^b	4	1	1	0	1	1

^aIn the other commercial assays category the following assay types are included: Line Probe assay, Screen +, Ampliquality HPV-type, HPV Screening real-time PCR, Alias: Harmonia HPV, Alias: Venus HPV, Cepheid Xpert HPV, NeoPlex, Sacace HPV screen and type, HPV Direct-Flow Chip, INNO-LiPA HPV and Acon Biotech High risk.

^bIn the other in-house assays category, the following assay types are included: multiplex qPCR, real-time PCR, MALDI-TOF, RFLP.



Table 6: Number of datasets that have detected all HPV types in optional samples.

HPV assay	Number of data sets	HPV types in panel			
		16 (1 IU/ µl)	18 (1 IU/ µl)	31/33/45/52/58 (100 GE/ µl)	35/39/51/56/59/68 (100 GE/ µl)
All assays	158	133	126	144	148
Abbott/Abbott Alinity m (Abbott)	6	6	6	6	6
Allplex HPV 28/Allplex HPV HR (Seegene)	2	2	2	2	2
Ampfire HPV screen	2	1	0	1	2
Anyplex II HPV28/Anyplex II HPV HR ^a (Seegene)	8	8	8	5	5
BD_Onclarity/Viper (BD)	7	7	3	7	7
Cobas4800/6800 (Roche)	12	12	9	12	12
Genotyping with Real-time PCR	21	21	21	21	21
Hybrid Capture 2 (Qiagen)	19	0	0	14	18
14 High-risk (HybriBio)	19	19	19	19	19
21 HPV GenoArray (HybriBio)	20	20	20	20	20
37 HPV GenoArray (HybriBio)	19	19	19	19	18
ScreenFire HPV RS	2	0	0	2	2
TypeSeq/TypeSeq2	5	5	5	5	5
Other commercial assays ^a	12	11	12	9	9
Other In-house assays ^b	4	2	2	2	2

^aIn the other commercial assays category, the following assay types are included: Line Probe assay, Screen +, Ampliquality HPV-type, HPV Screening real-time PCR, Alias: Harmonia HPV, Alias: Venus HPV, Cepheid Xpert HPV, NeoPlex, Sacace HPV screen and type, HPV Direct-Flow Chip, INNO-LiPA HPV and Acon Biotech High risk.

^bIn the other in-house assays category, the following assay types are included: multiplex qPCR, real-time PCR, MALDI-TOF, RFLP.

4.3 Results by sample number

The numbers of laboratories typing for and reporting correct HPV type, with **no false positive** HPV type detected are shown in table 7. The number of datasets without false positive results differs for each sample.



Table 7:

Proficiency of each sample. To achieve 100% proficiency full detection of HPV types 16 and 18 with 10 IU/ µl, HPV types 31, 33, 45, 52 and 58 with 1000 genome equivalents/ µl and no false positivity is required.

HPV types	HPV genome equivalents per 5 µl	Percent correct data sets (N)
16	1 IU	99.37 (157/158)
16	10 IU	86.08 (136/158)
18	1 IU	99.37 (157/158)
18	10 IU	86.08 (136/158)
31	1000 GE	93.67 (148/158)
33	1000 GE	98.10 (155/158)
45	1000 GE	93.04 (147/158)
52	1000 GE	96.20 (152/158)
58	1000 GE	97.47 (154/158)
31/33/45/52/58	100 ^a GE	98.73 (156/158)
35/39/51/56/59/68	100 ^a GE	98.10 (155/158)
Negative (TE buffer with 10 ng/µl human placenta DNA)	0	99.37 (157/158)

^a100 copies of each HPV type. IU: international units, GE: genome equivalents

4.4 Analysis of false positive results

To be considered as proficient for HPV testing no false positive result was accepted. The number of false positive samples by assay is shown in Table 8.

Table 8: Number of false positive HPV types detected per data set reported by assay used.

HPV assay	Number of data sets	No. of false positive samples per data set			
		0 samples	1 sample	2 samples	> 3 samples
All assays	158	144	11	1	2
Abbott/Abbott Alinity m (Abbott)	6	6	0	0	0
Allplex HPV 28/Allplex HPV HR (Seegene)	2	1	1	0	0
Ampfire HPV screen	2	2	0	0	0
Anyplex II HPV28/Anyplex II HPV HR ^a (Seegene)	8	5	2	1	0
BD_Onclarity/Viper (BD)	7	7	0	0	0
Cobas4800/6800 (Roche)	12	12	0	0	0



Genotyping with Real-time PCR	21	21	0	0	0
Hybrid Capture 2 (Qiagen)	19	19	0	0	0
14 High-risk (HybriBio)	19	19	0	0	0
21 HPV GenoArray (HybriBio)	20	20	0	0	0
37 HPV GenoArray (HybriBio)	19	18	1	0	0
ScreenFire HPV RS	2	0	2	0	0
TypeSeq/TypeSeq2	5	3	2	0	0
Other commercial assays ^a	12	10	1	0	1
Other In-house assays ^b	4	1	2	0	1

^aIn the other commercial assays category, the following assay types are included: Line Probe assay, Screen +, Ampliquality HPV-type, HPV Screening real-time PCR, Alias: Harmonia HPV, Alias: Venus HPV, Cepheid Xpert HPV, NeoPlex, Sacace HPV screen and type, HPV Direct-Flow Chip, INNO-LiPA HPV and Acon Biotech High risk.

^bIn the other in-house assays category, the following assay types are included: multiplex qPCR, real-time PCR, MALDI-TOF, RFLP.

We searched the data sets for patterns of consistent false positivity for any specific sample in the panel. The false positivities appeared to be essentially randomly distributed among the samples, indicating that the problem with false positives is usually not related to a property of the assays itself (e.g. cross-reactivity), but rather with the laboratory conditions of use (e.g. contamination).



5. Discussion

A total of 84 laboratories applied for participation in the 2022 screening proficiency study. The 2022 proficiency panels were distributed to all solicitants and a total of 158 datasets from 84 laboratories were returned before the deadline (January 2023).

We here report the results of the first international HPV screening proficiency study providing an overview of the performance of laboratories using various HPV testing methodologies in the screening setting. We have found that the majority of laboratories were highly competent, providing accurate results according to the preliminary proficiency criteria.

The main strengths of this study are that i) the content of the proficiency panel is traceable to international standards and ii) the wide participation from many laboratories makes it internationally comparable. The datasets submitted have been obtained from a variety of tests performed in many different settings, thus providing insights into the robustness of test performance when used in different laboratories.

Nevertheless, it seems that continued research is required in particular to define the optimal type of composition and analytical sensitivity for optimal clinical specificity.

A major limitation is that the proficiency criteria established in this study were provisional. Indeed, the results indicate that there is a need for additional research regarding the required analytical sensitivity, in particular for HPV16 and 18. We provisionally used the same requirement as for HPV genotyping proficiency studies (10 IU/ μ l) which was agreed on at an international WHO consensus workshop 2009. However, the fact that none of the 19 laboratories using the Hybrid Capture 2 (HC2) assay were able to detect HPV16/18 in this amount implies that the proficiency criteria for these two HPV types, HPV16 and 18, may need to be reconsidered. HC2 assay has been evaluated in several classic randomized controlled trials and is FDA approved. The reported sensitivity of the assay (detection limit) is 100 copies. Hortlund et al. found that only 3/49 (6%) HPV18 positive (HPV16 negative) invasive cancers that were preceded by viral amounts lower than 100 IU/ μ l, implying that maybe this analytical sensitivity (100 IU/ μ l) is sufficient. On the other hand, almost all assays used worldwide were able to detect HPV16/18 at 1 IU/ μ l.

The lack of established requirements for specificity was reflected in the results of this study. Particularly, this was clearly seen in the results obtained from the samples contained HPV types of lower oncogenicity (not HPV 16/18) at low viral loads, among which almost no cases of cervical cancer are found. and highest false positivity (5.75%) found for HPV35/39/51/56/58/59/66/68 in the



reference screening. As global HPV screening programs are implemented, a lack of consideration for specificity of the test may result in very large numbers of women unnecessarily being labelled as screen-positive.

Some of the HPV assays performed better than others, in agreement with what has already been reported in previous HPV DNA genotyping proficiency reports. However, the high proportion of datasets that were fully proficient already in this first round of proficiency test for HPV screening implies that the problems are limited and that a continued HPV screening proficiency program will readily be able to arrive at quite reliable performance of the global HPV screening laboratories.

In conclusion, a global proficiency study of HPV testing services for cervical screening programs has found that, although the results were mostly satisfactory, continued research on the required analytical sensitivity for optimal clinical sensitivity and clinical specificity is needed.



6. Conclusions and recommendations

This technical report summarizes the results obtained from the 1st HPV LabNet HPV DNA proficiency study targeting HPV screening assays. The study was open for participation to all laboratories across the globe.

The proficiency panel provided the possibility to evaluate the specificity and sensitivity for different HPV typing assays to correctly identify 13 high risk HPV types and 2 low risk HPV types, the HPV types that are the most important for HPV vaccine research as well as for HPV surveillance and monitoring.

We report that the global proficiency in HPV genotyping services had increased in the 2022 proficiency study. The increased proficiency is seen both in laboratories that had participated in previous studies as well as in laboratories participating for the first time. An increase in submitted datasets from the WPRO (102 datasets) was seen, surpassing the EURO region (38 datasets) which usually was the region with more laboratories participating in the previous proficiency studies.

The 2022 Global HPV LabNet HPV DNA proficiency panel further supports that it is possible to perform global studies comparing the sensitivity and specificity of different HPV typing assays, as well as the performance of participating laboratories, in a consistent manner that allows comparison of results generated by different laboratories worldwide and over time. Comparing the results from 2008, 2010, 2011, 2013, 2014, 2017, 2019, 2021 and 2022 Global HPV DNA PPs, we can see overall improvements.

Moreover, the highest overall proficiency (77% of datasets) was achieved in this proficiency study, 2022 – higher than in any previous proficiency study. We suggest that recommendations for HPV laboratory testing should continue to include a strong emphasis on the use of negative controls in the assays.

It is possible to achieve a global improvement in proficiency of HPV genotyping services. The improvement in proficiency seen in the 2021 and 2022 proficiency study suggests that continuing proficiency testing is helpful to sustain and improve accuracy and to avoid a deterioration in



proficiency, as seen in the last proficiency study in 2019. In the efforts to eliminate cervical cancer, the International HPV Reference Center will continue to issue PP yearly to promote proficiency in HPV testing services. Starting in 2022, we did also issue HPV screening panels (with screening-relevant concentrations of the HPV genotypes important for screening) to promote proficiency in HPV screening services as well.



Annex 1:

Call for participation: HPV DNA screening proficiency study 2022 (597)

External quality assessment for laboratories that perform **HPV screening**.

Testing for oncogenic HPV in cervical cancer screening is a globally recommended health policy. For HPV screening, it is important to detect the most carcinogenic viruses (HPV16 and 18) at high sensitivity, but for less carcinogenic viruses it is not important to determine the exact type.

Complete HPV genotyping is particularly important for vaccinology, but as the HPV genotyping proficiency panel contains 44 challenge samples, many laboratories have asked for a smaller proficiency panel tailored to what is important for assessing the quality of HPV screening services.

Participant laboratories will be asked to perform HPV screening using one or more of their usual HPV screening assays on the 12 challenge samples in this panel. The panel will evaluate the detection of the major oncogenic HPV types as well as detection of other oncogenic HPV types in groups.

Composition of sample material

The panel consists of 12 samples containing purified whole genomic plasmids of **HPV 16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, 59, 68a** and **68b** in a background of human cellular DNA. Each sample may contain either no HPV, a single HPV type or a mixture of HPV types at varying concentrations.

Participation fee

The price for one panel (12 samples as described above) is 330 Euros for commercial entities and 140 Euros for academic entities. Participants from low and lower middle-income countries (World Bank classification with GNI (gross national income) per capita: <4045 USD) can apply for waiving of fee. Laboratories that have outstanding payments from past Global HPV LabNet proficiency studies will need to clear their debts before their registration is accepted.

Data submission

The International HPV Reference Laboratory in Sweden is organizing this study in collaboration with the Swedish external quality assurance provider Equalis AB, who is responsible for management and distribution. Laboratories that have more than one assay are encouraged to provide results on each assay they commonly use. Data submitted will become the property of the organizers and may be analyzed for publication by the Global HPV LabNet either as an internal document or peer reviewed manuscript. All results will be handled in a coded anonymous fashion, with summaries grouped by method. The code linking data to originating laboratories will be kept confidential. Laboratories that provide data within the required time-frame will receive a copy of their own results and the summary data.

Scientific issues

Dr. Joakim Dillner
HPV International Reference Laboratory, Sweden



EQUALIS

Center for Cervical Cancer Prevention

E-mail: joakim.dillner@ki.se or www.hpvcenter.se

Registration

Application forms for the 2022 study can be found at: equalis.se

Preliminary dates

15th of June 2022: Registration for participation opens.

26th of August 2022: Registration for participation closes.

September/October 2022: Dispatch of panels begins.

Participation, management and practical issues

Equalis AB, Sweden www.equalis.se

E-mail: HPV@equalis.se

Annex 2:**Application for participating in
The 2022 Global HPV LabNet DNA Screening Study**

Fee for commercial entities: 330 Euros, Fee for academic entities: 140 Euros. Participants from low and lower middle-income countries (World Bank classification with GNI (gross national income) per capita: <3 975 USD) can apply for waiving of fee.	
Delivery address:	
Lab ID (from previous participation):	
Department /Laboratory	
Address	
City	Postal code
Province /State	Country
E-mail	Fax
Phone	
Invoice address (if different from above):	
VAT number:	
Department /Laboratory	
Address	
City	Postal code
Province /State	Country
Mode of payment (please check the preferred choice):	
By Credit Card/PayPal:	
By Invoice:	
Principal Investigator:	
First Name	
Surname (Title)	
HPV DNA typing experience in your laboratory	
Methodology used (may be more than one)	
Annual number of HPV typing tests performed	
Brief description of involvement in HPV surveillance or HPV vaccine development	

Return registration form by email or fax to: info@equalis.se. Fax: +46 18 69 31 46

Annex 3:**Source of panel material.**

Complete genomes of HPV cloned into plasmid vectors had been provided to the Lund University by the respective proprietors with a written approval to be used in this proficiency panel: Dr Ethel-Michele de Villiers (HPV types 16, 18 and 45), Dr Gérard Orth (HPV types 33, 39 and 68a prototype), Dr Elisabeth Schwarz (HPV 68b), Dr Saul Silverstein (HPV type 51), Dr Attila Lörincz (HPV types 31, 35 and 56), Dr Wayne Lancaster (HPV type 52) and Dr Toshihiko Matsukura (HPV types 58 and 59).

The HPV genomes are cloned into different cloning vectors: HPV 16 in pBR322 at position 6152, HPV 18 in pGEM-5Zf vector in the L2 region, HPV 31 in pT713 at position 3362, HPV 33 in pBR322 at position 2797, HPV 35 are cloned in two fragments 5012-956 and 956-5012 in pT713, HPV 39 in pGEM4z at position 1714, HPV 45 in pGEM4Z at position 75, HPV 51 in pGEM4z at position 4511, HPV 52 in pUC19 at position 7559, HPV 56 in pT713 at position 5521, HPV 58 in pGEM4Z at position 1158, HPV 59 in pUC9 at position 69, HPV 68a prototype in a bluescript vector, and HPV68b (ME180) of about 7 kb containing L1, L2, E1, E2, E4, E5, E6, E7 with an incomplete E2 gene in pCR4-TOPO.

The nucleic acid sequences for each of these HPV genomes have been reported previously and are available in Gene Bank with the following accession numbers; HPV 16 nr K02718; HPV 18 nr X05015; HPV 31 nr J04353; HPV 33 nr M12732; HPV 35 nr M74117; HPV 39 nr M62849; HPV 45 nr X74479; HPV 51 nr M62877; HPV 52 nr X74481; HPV 56 nr X74483; HPV 58 nr D90400; HPV 59 nr X77858; HPV 68a nr X67161 and HPV 68b nr FR751039.

Annex 4:

**List of participants in the HPV LabNet Proficiency Study
of HPV DNA Screening, 2022**

Laboratory details
AMR/PAHO – Regional Office for the Americas (27labs)
<p>Maria Alejandra Picconi Oncogenic Viruses Service National Reference Laboratory for Papillomavirus National Institute of Infectious Diseases- ANLIS "Dr. Malbran" Av. Velez Sarsfield 563 C1282AFF- Buenos Aires ARGENTINA Email: mapicconi@anlis.gov.ar; mapicconi@gmail.com</p>
<p>Irene Sorin Hospital Eva Perón General San Martín Provincia de Buenos Aires ARGENTINA</p>
<p>Susana Ageitos / Patricia Amarante Hospital Mi Pueblo Florencio Varela Provincia de Buenos Aires ARGENTINA</p>
<p>Monica Fazzito Ipar Hospital Paroissien La Matanza Provincia de Buenos Aires ARGENTINA</p>
<p>Alicia Morel Hospital Vidal Corrientes ARGENTINA</p>
<p>Patricia Alejandra Perez / Cintia Debora Bravo Hospital San Martin Paraná Entre Ríos ARGENTINA</p>

Karina Ferreira / Sebastian Leiva Dirección Provincial de Laboratorios Formosa ARGENTINA
Oscar Marin Hospital Pablo Soria Jujuy ARGENTINA
Hugo Gongora Jara / Noelia Carreño Hospital Regional Dr. Enrique Vera Barros La Rioja ARGENTINA
Dra Gabriela Andino Hospital Castro Rendon Neuquén ARGENTINA
Silia Chaves / Ana Claudia Gonzalez Hospital Dr. Pedro Moguillansky Cipolletti Río Negro ARGENTINA
Siliva Yocca Hospital San Bernardo Salta ARGENTINA
Agustina Lacaze Lab, Salud Pública Dr. Dalmiro Pérez Laborda San Luis ARGENTINA
Ana Viñuales Instituto de Maternidad Nuestra Señora de las Mercedes Tucumán ARGENTINA
Gabriel Jeanney Hospital Oncológico J. B. Iturraspe Santa Fe ARGENTINA

Facunto Tomas Sachis / Eliana Georgef Horvat Hospital Perrando Resistencia Chaco ARGENTINA
Julieta Guyot Hospital Prof A. Posadas Haedo Provincia de Buenos Aires ARGENTINA
Viviana Chequer Hospital Regional Dr. Ramón Carillo Santiago del Estero ARGENTINA
Sandra Robledo Hospital San Juan Bautista Catamarca ARGENTINA
Staciuk Federico Javier Hospital Escuela de Agudos Dr. Ramón Madariaga Posadas Misiones ARGENTINA
Diego Chouhy Piso 3 de la Nave 4 del Campus Corporativo Cooperativo Ruta Provincial N 16 km 5 Interseccion autopista Rosario-BsAs y AO12 Alvear ARGENTINA Email: dchouhy@detxmol.com.ar
Juliana Comerlato Laboratório Epiclin - Hospital Moinhos de Vento Rua Ramiro Barcelos 630, sala 817 BR-90035 -0 PORTO ALEGRE BRAZIL Email: juliana.comerlato@hmv.org.br
Ruby Aguillon Central Medical Laboratory Ministry of Health BELIZE Email: celsocarrera@chr.health.gov.bz ; rbyaguillon@chr.health.gov.bz

Compton Beecher Caribbean Genetics FMSTR Complex, Block B, Level4 University of the West Indies Kingston JAMAICA Email: COMPTON.BEECHER@UWIMONA.EDU.JM
Casey Dagnall Cancer Genomics Research Laboratory National Cancer Institute 9615 Medical Drive US-20850 Rockville USA Email: dagnallc@mail.nih.gov
Guillermo Rodriguez Comision honoraria de Lucha contra el Cancer Brandzen 1961 of 1201 11200 Montevideo URUGUAY Email: lauraga@adinet.com.uy
Maria Liz Bobadilla Ministerio de Salud Publica y Bienstar Social Avenida Venezuela y Teniente E 59521 Asuncion PARAGUAY Email: bobadillaml@gmail.com ; inmunologia.lcsp@mspbs.gov.py
Carolina Teran Real y Pontificia de San Francisco Xavier de Chuquisaca Faculty of Med.Univ.Mayor Rene Moreno Str Sucre BOLIVIA Email:
EURO –European Region (30 labs)
Gro Kummeneje Presthus Akershus University Hospital Avdelning for mikrobiologi og smittevern HPV Reference Laboratory Sykehusveien 25 1478 Lørenskog NORWAY Email: gro.kummeneje.presthus@ahus.no

<p>Christine Penz Akershus University Hospital Department of Pathology Sykehusveien 25 1478 Lörenskog NORWAY Email: Christine.penz@ahus.no</p>
<p>Inger-Lise Haakensen Oestfold County Hospital Dep. Of Gene Technology and Infectious Serology Postbox 300 1714 GRAALUM NORWAY Email: Inghak@so-hf.no</p>
<p>Anna Veronica Reber Norén AND Alba Kaci Oestfold Hospital Trust Center for Laboratory Medicine, Department of Gene technology Kalnesveien 300 1714 Grålum NORWAY Email: annano@so-hf.no Email: albkac@so-hf.no</p>
<p>Grete Moås Oslo University Hospital, Radiumhospital Department of Pathology Ullernchaussen 56 0379 Oslo NORWAY Email: grmoaa@ous-hf.no</p>
<p>Nordlandssykehuset Department of Pathology 8092 Bodø NORWAY</p>
<p>Gudrun Hovstein, Siri Borchgrevink-Persen St. Olav hospital Department of pathology Postbox 3250 Torgarden 7006 Trondheim NORWAY Email: gudrun.hovstein@stolav.no Email: siri.borchgrevink-persen@stolav.no</p>
<p>Undis Venneström Stavanger universitetssykehus Avd for Patologi, Cytologi lab Gerd Ragna Block Thorsens gate 8 4011 Stavanger</p>

<p>NORWAY Email: undis.elisabeth.vennestrom@sus.no</p>
<p>Katrine Eklo Universitetssykehuset i Nord-Norge HF Klinisk Patologi Seksjon for spesialanalyser Sykehusvegen 38 9019 Tromsø NORWAY Email: Katrine.eklo@unn.no</p>
<p>Raja Levorsen Patologiavdelingen Sykehuset i Vestfold HF Halfdan Wilhelmsens alle 17 3103 Tønsberg NORWAY Email: raja.levorsen@siv.no</p>
<p>Christine Dønheim Ålesund sjukehus Avdeling for patologi 6026 Ålesund NORWAY Email: Christine.Holtvedt.Donheim@helse-mr.no</p>
<p>Oda Teige Kalsås Haukeland universitetssjukehus Avdeling for patologi Postboks 1400 5021 Bergen NORWAY Email: oda.teige.kalsas@helse-bergen.no</p>
<p>Sarah Carlsson Centrum för cervixcancerprevention, F56 Karolinska Universitetssjukhuset Huddinge 141 20 Huddinge SWEDEN Email: sarah.carlsson@regionstockholm.se</p>
<p>Fanny Huynh Söderberg ABC Labs/QC_QA Management Tomtebodavägen 16 17165 Solna SWEDEN Email: fanny@abclabs.se</p>
<p>Anna Skoglund SYNLAB Medilab Nytorpsvägen 30</p>

183 53 Täby SWEDEN Email: anna.skoglund@synlab.se
Karin Elfving Avd för klinisk mikrobiologi Falun lasarett 791 82 Falun SWEDEN Email: karin.elfving@regiondalarna.se
Pernilla Tjernberg Lab.medicin Västernorrland Länssjukhuset S-vall/H-sand 851 86 Sundsvall SWEDEN Email: pernilla.tjernberg@rvn.se
Kate Cuschieri Scottish HPV Reference Laboratory (SHPVRL) Specialist Virology Centre Dept of Laboratory medicine Royal Infirmary of Edinburgh 51 Little France Crescent Edinburgh EH16 4SA, UK Email: Kate.Cuschieri@nhslothian.scot.nhs.uk Kirsty.E.Gray@nhslothian.scot.nhs.uk
Elisabetta Tanzi Department of Biomedical sciences for Health Via C. Pascal 36 201 33 Milan ITALY Email: elisabetta.tanzi@unimi.it
Jesper Bonde, Helle Pedersen Molecular Pathology Laboratory Dept. Pathology, afs 134 Hvidovre Hospital Kettegård Alle 30 26 50 Hvidovre DENMARK Email: Helle.Pedersen.03@regionh.dk
Rosemarie Preyer GenID GmbH Ebinger Strasse 4 DE-72479 Strassberg GERMANY Email: preyer@aid-diagnostika.com

Thomas Iftner Institute for Virology Universityhospital Tuebingen Elfriede-Aulhorn-Strasse 6 DE-72074 Tübingen GERMANY Email: angelika.iftner@med.uni-tuebingen.de
Steffi Silling, Ulrike Wieland Institute of Virology, University of Cologne Fuerst-Pueckler-Strasse 56 50935 Koeln GERMANY Email: steffi.silling@uk-koeln.de
Martin Daeumer Institute for Immunology and Genetics Pfaffplatz 10 67655 Kaiserslauten GERMANY
Anna Eis-Huebinger Institute of Virology University Hospital Bonn Venusberg-Campus 53127 Bonn GERMANY
Alexander Adam Institute of Pathology University Hospital Cologne and MVZ Oberberg und Siegen Wilhelm-Breckow-Allee 20 51643 Gummersbach GERMANY
Andreas Kaufmann Clinic for Gynecology Charité Campus Virchow Clinic Suedring 11 (HPV Laboratory) Augustenburger Platz 1 133 53 Berlin GERMANY
Sonia Perez Castro Servizo Microbioloxia Hospital Meixoeiro, CHUVI 36214 Vigo SPAIN Email: sonia.maria.perez.castro@sergas.es
Asuncion Olma-Sevilla Vitro S.L P. T. Ciencias Salud

<p>Avda Conocimiento 100 18016 Granada SPAIN Email: asuncion.olmo@vitro.bio; belen.espigares@vitro.bio</p>
<p>Jean-Luc Pr��tet CHRU Besancon J. Minjoz Niv 2 Magasin Cancero Biologi Rue Charles Bried Fr-25030 Besancon FRANCE Email: lpepin@chu-besancon.fr</p>
<p>Yalin Kilic Izmir University of Economics Faculty of Engineering Department of Biomedical Engineering TR-35330 Izmir TURKEY Email: kilic.yalin@ieu.edu.tr</p>
<p>SEARO –South-East Asia Region (2 labs)</p>
<p>Sompong Vongpunsawad Dept. Pediatrics, Faculty of Medicine Chulalongkorn University 1873 Rama 4 Road Pathumwan, BANGKOK 10330 THAILAND Email: sompong.vo@chula.ac.th</p>
<p>Neeraj Jain Room No 111, Research Laboratory IV (Biological Sciences) CHARUSAT, Changa Charotar Univ. of Scines&Techno P D Patel Inst of appl.science IN-388421 ANAND INDIA Email: neerajjain.as@charusat.ac.in</p>
<p>WPRO –Western Pacific Region (27 labs)</p>
<p>Joohee Lee Genematrix 700 Daewangpangyo-ro Bundang Korea Bio Park B Bldg. 8F KR-13488 Seongnam-si SOUTH KOREA</p>

Email: mangoid@genematrix.net
Yaeseul Suh Seegene Inc 91 Ogeum-ro, Songpa-gu KR-05548 Seoul SOUTH KOREA Email: yssuh@seegene.com
Xiaoyun Li The first affiliated hospital of Anhui medical university 230000 Hefei CHINA Email: xiaoyunli@hybribo.cn
Wenjing Hu Jinan Hybribo Medical laboratory 999077 Hong kong CHINA Email: wjhu@hybribo.cn
Liejun Li Molecular Biology Laboratory 27/F Bonham TradeCenter Hongkong 50 Bonham Strand, Sheung Wan 521000 Chaozhou CHINA Email: ljli@hybribo.cn
Ting Wang Shenzhen Hybribo Medical Laboratory No.1301-76 Guanguang Road Room A401, Building 1, Yinxing Zhijie Phase II Xinlan Community, Guanlan Street 518110 Shenzhen CHINA Email: tingwang@hybribo.cn
Xu Feng Guangzhou Hybribo Medical Laboratory No 3 Sheung Yue Road Kowloon Bay 35/F Enterprise Square Two CN-510555 GUANGZHOU CHINA Email: fengxu@hybribo.cn
Walter Zhang Shanghai ZJ Bio-Tech Co., Ltd. Lab Building 26, 588 Xinjunhuan Road 2011 14 Shanghai

CHINA Email: walter_zhang@liferiver.com.cn
Lala Du Hangzhou HybriBio Medical Laboratory No.448-1, North Donghu Road Room 301, Building 32 311100 Hangzhou CHINA Email: lidu@hybriBio.cn
Xiaoying Liu The First hospital of Yulin High-tech Zone, Yulin city No.93 Yuxi Avenue 719000 Yulin CHINA Email: 769836113@qq.com
Li Tian Pathology of Tangshan Central Hospital 601-1 Changning Road Lubei District 63008 Tangshan CHINA Email: 305726988@qq.com
Weiyang Liu Foshan Nanhai HybriBio Medical Laboratory No. 14-2 Guihe Road, Lishui T Nanhai District, Guangdong 528200 Foshan CHINA Email: wylu@hybriBio.cn
Tao Zou Nanchang HybriBio Medical Laboratory No. 188 Longtan Road Qingshanhu District, Jiangxi Province 330000 Nanchang CHINA Email: tzou@hybriBio.cn
Joanne Zhou ACON Biotech (Hangzhou) Co., LTD West Lake District No.210 Zhenzhong Road 310030 Hangzhou CHINA Email: joanne.zhou@aconlab.com.cn
Xiang Liu

No 9. Weidi Road Qixia District Nanjing HybriBio Medical Laboratory 5 th Floor, Building D6 CN-210000 Nanjing CHINA Email: xliu@hybriBio.cn
Jinhui Shao Xian HybriBio Medical Laboratory, Room 10101-10106 Building B7, caotan 10 th Road China Electronics Xian Industry 710000 Xian CHINA Email: yudou@hybriBio.cn
LiPing Zhang Zhengzhou HybriBio medical Laboratory CN-450000 Zhengzhou CHINA Email: lpzhang@hybriBio.cn
Ying Yin Zhang Wuhan HybriBio Medical Laboratory Co. Ltd CN-430000 Wuhan CHINA Email: yyzhang@hybriBio.cn
Juanjuan Zhu Shanghai HybriBio Medical Laboratory CN-201318 Shanghai CHINA Email: jjzhu@hybriBio.cn
Yuying Lv Beijing HybriBio Medical Laboratory Ltd. CN-102206 Beijing CHINA Email: yylv@hybriBio.cn
Minfang Li Changsha HybriBio Medical Laboratory Ltd. CN-410000 Changsha CHINA Email: minfangli@hybriBio.cn
Lu Yuchen Yaneng Bioscience Co Ltd Gaoxingqi Industrial Park, 10 th Floor, Building 1, Phase 2 Liuxian 1 st Road, Baoan District 518000 Shenzhen CHINA

Email: chenziyu@yanengbio.com
Canbin Zhuo Shantou Kaipu medical laboratory Co Ltd CN-515041 Shantou CHINA Email: cbzhuo@hybribo.cn
Xinle Feng Chaozhou Hybribo medical Laboratory Co Ltd CN-521000 Chaozhou CHINA Email: fengxl@hybribo.cn
Taiyangsheng Yin Shenyang Hybribo Medical Laboratory Co Ltd CN-110000 Shenyang CHINA Email: ytys@hybribo.cn
Shixian Geng Kunming Hybribo Medical Laboratory Ltd CN-650217 Kunming CHINA Email: sxgeng@hybribo.cn
Yijing Zeng Chengdu Hybribo Medical Laboratory Ltd Chengdu CHINA Email: yijingzeng@hybribo.cn